

CALIBRATION CERTIFICATE

Certificate No	QU-2060
----------------	---------

Contact Name:	Andrew Bush
Customer Name:	M Gaze & Co. Ltd
Site Address:	Thurlton Norwich NR14 6NZ
Email:	Andrew@gaze.co.uk


All Test Weights / Test Rigs used are traceable to National Standards.

LINEARITY		
Applied Weight	Reading As Found	Reading As Left (Definitive)
0 kg	N/A	0 kg
400 kg	New Weighbridge	400 kg
10000 kg		10000 kg
20000 kg		20000 kg
30000 kg		30000 kg
44020 kg		44020 kg

EQUIPMENT DETAILS	
Make / Model:	Dini Argeo 3590 ETD
Serial Number:	93932665
Customer ID No.	N/A
Location:	Site
Capacity:	50,000 kg
Division Size:	20 kg

CALIBRATION DETAILS	
Accuracy:	+/- 0.1%
Test Weights Certificate Number(s)	TR10905, TJ0005
Calibration Date:	06 February 2020
Calibration Due:	February 2021

ECCENTRIC LOADING		
Corner 1	N/A	10000 kg
Corner 2	New Weighbridge	10000 kg
Corner 3		10000 kg
Corner 4		10000 kg

CERTIFICATE ISSUED BY	
Name:	Ellis Blyth
Signature:	
Date:	11 February 2020

QF1 – Revision 3